

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2009

| | | | | | | |
|--|---------------------|---------------------------|---------|----------------------|-----------|---------|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | 130129 | ✓ | | |
| Name | Jean Achee | ID# | 1470 | Date | 6/10/2009 | ✓ |
| A | Agency | Wasilla Police Department | Phone # | 907-352-5401 | | |
| | Instrument Location | Wasilla Police Department | | | | ✓ |
| B | Alco S/N | X124625 | ✓ | Target Value | .081 | ✓ |
| | | | | High Pressure | 800 | |
| | Alco Test Values | .082 | ✓ | .082 | ✓ | |
| | | 1 st Alco | | 2 nd Alco | | |
| | Signature | Jean Achee | ✓ | | | |
| | | | | | | 6/25/09 |
| (OVER) | | | | | | |

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 20th day of June 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

JUNE 10, 2009 ✓

OPERATOR'S NAME: ✓
ACHEE, JEAN P. *COB 6/25/09*
OPERATOR'S NUMBER: 1470
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567890
DEPT/AGENCY: WAS1
CASE/REPORT: 1234567890
TEST TYPE: V
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X124625 ✓

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .081 ADJUSTED FOR 29.75 in | ✓ | 14:21 |
| ALCO TARGET | .080 | 14:22 |
| BLANK TEST | .000 | 14:22 |
| INTERNAL STANDARD | VERIFIED | 14:22 |
| ALCO TV 29.77 in | .082 | 14:22 |
| BLANK TEST | .000 | 14:23 |
| SUBJECT SAMPLE | .000 | 14:24 |
| BLANK TEST | .000 | 14:25 |
| ALCO TV 29.75 in | .082 | 14:25 |
| BLANK TEST | .000 | 14:26 |

COB 6/25/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

JUNE 10, 2009 ✓
TIME 14:27 *80*

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 01/08/09 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.75 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~`